

## **Pewatron Application Note PAN-03: Zirconia oxygen gas sensor – cross sensitivity to other gases and materials**

The FCX zirconia oxygen sensor works excellent and reliable in a number of applications, including CO<sub>2</sub> incubators, nitrogen and oxygen concentrators, oxygen transmitters and in fruit & vegetable storage applications. Zirconium oxide (with proper dopant elements; zirconia) when heated > 350 °C is penetrable for oxygen ions. The FCX oxygen series has the working point at 450°C. A voltage applied to the oxygen-sensing element, pumps the oxygen out of a closed inner chamber. At a constant gas pressure, the quantity of oxygen pumped out is equal to the quantity of oxygen molecules diffusing in through a small capillary hole. It is independent of the voltage applied between the electrodes within a certain range. The measurement current is proportional to the quantity of oxygen molecules pumped away. Because of the high temperature and the voltage driving force for oxygen diffusion, there is number of gases and materials, which can result in a false oxygen concentration measurement or can destroy the oxygen sensing element. In this application note, we will discuss most of these gases and material that have an adverse effect on the oxygen-sensing principle of the zirconia sensor.

### **Combustible Gases**

Small amounts of combustible gases does not burn at the hot sensor surfaces, but they can decompose because of a high applied voltage across the sensor element. Below the decomposition voltage, the sensor will measure the residual oxygen pressure.

Gases investigated:

- H<sub>2</sub> (Hydrogen) up to 3%
- CO (Carbon Monoxide) up to 3%
- CH<sub>4</sub> (Methane) up to 4%
- CO<sub>2</sub> (Carbon dioxide) up to 80%

### **Heavy Metals Vapors**

Pt (Platinum), Zn (Zinc), Cd (Cadmium), Pb (Lead), Bi (Bismuth) will have a negative effect on the catalytic properties of the oxygen sensing electrodes. Exposures to such metal vapors must be avoided.

### **Halogen and Sulphur Compounds**

Small amounts (< 100ppm) of Halogens and/or Sulphur compounds have a negative effect on the lifetime and performance of the oxygen sensor. These types of gases cause sensor degradation.

Gases investigated:

- Halogens, F<sub>2</sub> (Fluorine), Cl<sub>2</sub> (Chlorine)
- HCL (Hydrogen Chloride), HF (Hydrogen Fluoride)
- SO<sub>2</sub> (Sulphur Dioxide), H<sub>2</sub>S (Hydrogen Sulphide)
- Freons, CFCs (Chlorofluorocarbon)

### **Nitrous gases**

Exposure to nitrous gases in ppm levels change the catalytic effect of the Pt-electrodes and result in false signals. This can be the case for NO<sub>x</sub> and NH<sub>x</sub> gases.

Gases investigated:

- NO<sub>2</sub> (Nitrogen Dioxide)
- NO<sub>x</sub> (Nitrogen oxides)
- NH<sub>3</sub> (Ammonia)

### **Others**

Gases evaporating from silicone compounds or rubbers are to be avoided. Ppm levels of SiO<sub>2</sub> and Si reacts with the Platinum electrodes or blocks the pores and active parts of the electrodes.

VOC or other organic gases in high concentrations and with a low decomposition temperature can have a negative influence on the lifetime and performance of the oxygen sensor

For humidity and water condensation issues concerning the sensor performance, please refer to Pewatron application note PAN-04

Headquarter Switzerland:  
Angst+Pfister Sensors and Power AG

Thurgauerstrasse 66  
CH-8050 Zurich  
Phone +41 44 877 35 00  
sensorsandpower@angst-pfister.com

Office Germany:  
Angst+Pfister Sensors and Power  
Deutschland GmbH  
Edisonstraße 16  
D-85716 Unterschleißheim  
Phone +49 89 374 288 87 0  
sensorsandpower.de@angst-pfister.com



## We are here for you. Addresses and Contacts.

### Sales Germany & Austria

Geometrical sensors  
Other products

Kurt Stritzelberger  
Phone +49 89 374 288 87 22  
kurt.stritzelberger@angst-pfister.com

Pressure sensors  
Other products

Gerhard Vetter  
Phone +49 89 374 288 87 26  
gerhard.vetter@angst-pfister.com

Gas sensors and modules

Peter Felder  
Phone +41 44 877 35 05  
peter.felder@angst-pfister.com

### Sales Switzerland & Liechtenstein

Postcode 3000 – 9999

Basil Frei  
Phone +41 44 877 35 18  
basil.frei@angst-pfister.com

Postcode 1000 – 2999

Christian Mohrenstecher  
Phone +41 76 444 57 93  
christian.mohrenstecher@angst-pfister.com

### Sales International Key Accounts

Peter Felder  
Phone +41 44 877 35 05  
peter.felder@angst-pfister.com

### Sales Other Countries / Product Management

Pressure Sensors  
Load Cells

Philipp Kistler  
Phone +41 44 877 35 03  
philipp.kistler@angst-pfister.com

Gas sensors  
Gas sensor modules

Dr. Thomas Clausen  
Phone +49 89 374 288 87 24  
thomas.clausen@angst-pfister.com

Flow / Level / Medical products

Dr. Adriano Pittarelli  
Phone +49 89 374 288 87 67  
adriano.pittarelli@angst-pfister.com

Power supplies

Sebastiano Leggio  
Phone +41 44 877 35 06  
sebastiano.leggio@angst-pfister.com

Linear position sensors  
Angle sensors

Eric Letsch  
Phone +41 44 877 35 14  
eric.letsch@angst-pfister.com

Accelerometers  
Sensor elements

Christoph Kleye  
Phone +49 89 374 288 87 61  
christoph.kleye@angst-pfister.com

Drive technology  
CH Postcode 5000 – 9999 / DE

Roman Homa  
Phone +41 76 444 00 86  
roman.homa@angst-pfister.com

Drive technology  
CH Postcode 1000 – 4999 / AT / IT / FR

Christian Mohrenstecher  
Phone +41 76 444 57 93  
christian.mohrenstecher@angst-pfister.com

Harald Thomas  
Phone +49 89 374 288 87 23  
harald.thomas@angst-pfister.com